



DECEMBER 11-12, 2025
۲۱ و ۲۰ آذر ماه ۱۴۰۴



دومین کنگره
پلاسما پزشکی ایران

The 2nd Congress on Plasma Medicine

دبیرخانه دائمی کنگره
پلاسما پزشکی ایران
www.plasmamedsym.ir



Experimental–Numerical Characterization of Cold Atmospheric Pressure Plasma Jet for Biomedical Applications Using Moiré Deflectometry

Erfan Ghane Sheykh Abadi^{1*}, Heydar Izdanshan¹, Mahmoud Dehghan², Fatemeh Jamal³, Mohammad Haqparast⁴, Lily Darvish⁴

¹Department of Optics and Laser Physics, Islamic Azad University, Shiraz, Marvdasht, Iran

²Department of Radiology, Faculty of Paramedicine, Hormozgan University of Medical Sciences, Hormozgan, Iran

³PhD student in Optics and Lasers, University of Zanjan, Zanjan, Iran

⁴Department of Radiology, Faculty of Paramedicine, Hormozgan University of Medical Sciences, Bandar Abbas, Iran

OPEN ACCESS

Citation:

Ghane Sheykh Abadi E, Izdanshan H, Dehghan M, Jamal F, Haqparast M, Darvish L. Experimental–Numerical Characterization of Cold Atmospheric Pressure Plasma Jet for Biomedical Applications Using Moiré Deflectometry. *Iran Biomed J. Supplementary* (2026): 66.

ABSTRACT

Introduction: Cold atmospheric pressure plasma jets (CAPJs) have shown significant potential for biomedical applications due to their non-thermal nature and high production of reactive species. In this study, we assessed the experimental-numerical characterization of CAPJs for using in biomedicine using Moiré deflectometry.

Materials and Methods: The electron density and refractive index distributions measured in the original Moiré deflectometry experiment were used as input data. A numerical model was implemented in MATLAB to reconstruct the continuous axial profiles $n_e(z)$ and $\Delta n(z)$ based on exponential fitting of the experimental points. The plasma was generated using a helium jet at atmospheric pressure, with an inner quartz tube of 1 mm diameter and a central stainless-steel electrode of 0.3 mm diameter. The applied voltage and gas flow rate were 5 kV-10.2 kV (18 kHz AC) and 0.5-2 L/min, respectively. Refractive index variation was calculated using the empirical relation is, $n-1=C_1 p/kT-C_2 N_e$, where $C_2=1.9292 \times 10^{-8}$ (cm³). The medically relevant parameters—plume length (for $n_e \geq 10^{18}$ m⁻³) and the RONS-proxy (integral of $n_e(z)$)—were then derived from the reconstructed profiles.

Results and Discussion: The numerical reconstruction successfully reproduced the experimental Moiré results, showing an exponential decay of n_e along the plasma axis. At 10.2 kV and 2 L/min, the maximum n_e near the nozzle was 3.1×10^{19} m⁻³, decreasing to 1.3×10^{19} m⁻³ at 16 mm. The effective plume length, defined as the distance where $n_e \geq 10^{18}$ m⁻³, was about 5.6 mm for high-power mode and 3.2 mm for low-power mode. The change in refractive index (Δn) derived from the Moiré analysis exhibited values of the order of 10^{-6} , confirming the optical sensitivity of the plasma region close to the nozzle. From a biomedical perspective, the obtained electron density range (10^{18} – 10^{19} m⁻³) is sufficient to produce reactive oxygen and nitrogen species (RONS) while maintaining gas temperatures below 50°C—safe for biological tissue exposure. Hence, the helium plasma jet characterized in the original experiment can be directly utilized for plasma medicine applications such as surface sterilization, cell stimulation, and wound treatment.

Conclusion: The combination of Moiré deflectometry and numerical reconstruction provides a powerful approach to link plasma diagnostics with biomedical functionality. The reconstructed axial profiles reveal that the plasma jet operates within the safe and effective range for biological applications. The model also offers a predictive tool to optimize voltage and gas flow parameters for future plasma medicine studies.



This article is licensed under a Creative Commons Attribution-NonDerivatives 4.0 International License.

Keywords: Biomedical application, Cold atmospheric plasma, Electron density, Moiré deflectometry, Refractive index

Corresponding Author: Erfan Ghane Sheykh Abadi

Department of Optics and Laser Physics, Islamic Azad University, Shiraz, Marvdasht, Iran; E-mail: saharanjarah1814@gmail.com



Iranian Biomedical Journal Supplementary (February 2026): 66